


Test Report

Personnel	SUN HWA JOO	Tel. No.	82-2-555-0500
Client	Affiliation MALTANI Corp.	E-mail	shj26@maltani.co.kr
	Address Maltani B/D, Yeoksam-Ro, Kangnam-Gu, Seoul 06196, Korea		
Request	Virucidal Activity Test by LED Irradiation		
Product	CLEAN SERIES MODULE		
Purpose of Use on the Product	Sterilization		
Test Virus	COVID-19 (SARS-CoV-2)	Cell Line	Vero E6
Test No.	KR-2009-042-MTN02-C	Test Period	2020.09.14-09.21
Reaction Time	60 min, 120 min	Titration	CPE
Test Temperature	Room Temperature (Approx. 20°C)	Tester	Hansam Cho 

Test Result

Product Name	Virus Titer TCID ₅₀	Distance	Treatment time	Virus Reduction Rate	
				(log)	(%)
CLEAN SERIES MODULE	6.81x10 ⁶	50 cm	60 min	1.58	97.34%
			120 min	2.08	99.16%

Result: As a result of the sterilization test for COVID-19 (SARS-CoV-2) by LED (405nm) generated in the CLEAN SERIES MODULE of MALTANI Corp., it showed 99.16% virucidal effect in 120 minutes treatment at a distance of 50cm.

October 05, 2020

Test Manager: Young Bong Kim (Seal)

KR BIOTECH Co., Ltd.



* This test report is a result limited to the sample and sample name provided by the client and does not guarantee the quality on the overall product.
* This report cannot be used for PR, advertising and litigation purposes, and use of this report other for its original purpose is prohibited.